## Resistance, potentiometer measurement

- Optional analogue output

- Mounting into panels with thickness up to 50 mm





- Plug in terminal with protection IP54
- Plug in terminal with protection IP65 (see reference)
- Protection IP65 in combination with analog output see PVE 4.xx6.6xx
- Reference: Decimal point, plus sign, blank digit have to be pretended!
- Analog output 0-10 VDC/10 mA
- Analog output 0-20 mA/load 500 Ω
- Analog output 4-20 mA/load 500 Ω
- Analog output 0-10 VDC/10 mA
- (power supply 24 VDC galvanically insulated)
- Analog output 0-20 mA/load 500  $\Omega$  (power supply 24 VDC galvanically insulated)
- Analog output 4-20 mA/load 500 Ω (power supply 24 VDC galvanically insulated)
- Analog output with customer specified offset

## The measuring inputs are not galvanically insulated from the analog output!

- Dimension strip selectable (7 characters max)
- Other power supplies on demand
- Setpoints see type PVE4.xx6.6xx

# Technical data, handling

Dimensions	Housing Assembly cut out Fastening Housing material Protective System Weight Connection	72 x 36 x 97 mm, including screw terminal $68.0^{+0.7}$ x $33.0^{+0.6}$ mm special quick plastic clamp proper to fix in wall thickness up to 50 mm PC/ABS-plastics blend, colour black, UL94V-0 at the front IP40 connection IP00 approx. 0.190 kg at the rear side via terminals up to 2.5 mm <sup>2</sup>
Input	Measuring range	1 KΩ - 10 KΩ 10 KΩ - 100 KΩ 100 KΩ - 1 MΩ offset adjustment supported by offset potentiometer
Output	Analogue output	0-10 VDC/10 mA (0.1 % of measuring value, +/-0.05 % of full scale) 0-20 mA, 4-20 mA - load 500 Ohm (0.1 % of measuring value, +/-0.05 % of full scale)
Accuracy	Resolution Nonlinearity Temp. drift Measuring principle	+/- 1999 digit +/-0.1 % of measuring value, +/- 1 digit 100 ppm/K Dual-Slope-Integration
Power Unit	Supply voltage Power consumption	230/115 VAC +/- 10 % (50-60 Hz), 24 VDC (18-30 V), 24 VDC +/-10 % galvanic insulated max. 5 VA
Indication	Display Measuring time Overflow Decimal point Blanking Plus-sign	LED with 7 segments, 14 mm high, red 3½-digit = indication 1999 1 second by showing "1" on the fourth digit adjustable by bridging on front side blanking out of last digit (selectable by bridge) selectable by bridging on front side
Ambient conditions	Working temperature Storing temperature	0 up to + $60^{\circ}$ C -20 up to + $80^{\circ}$ C

Housing:



<u>CE-sign</u> For unlimited use of the instrument within the directives for electromagnetic compatibility 89/336/EC analogue input wires have to be used with shielded cable and cable's shield connected to earth ground at one end only.

### Setting

- 1. Connect the instrument according to the wiring diagram and turn power on.
- 2. Setting of indication value: Remove the front pane using the eject gap.
- 3. Set the resistance value and adjust the desired indication value by means of the potentiometer.
- 4. In order to achieve maximum value indication of 1999, the following minimum resistance values are required at the various measuring inputs:

Measuring input	1 MΩ	100 KΩ	10 KΩ
Resistance (min)	500 KΩ	50 KΩ	5 KΩ